


1	Aspartic acid(Asn) 0.6 pmol	2	Gltamic acid(Gln) 0.6 pmol	3	Serine(Ser) 0.6 pmol
4	Glycine (Gly) 0.6 pmol	5	Histidine(His) 0.6 pmol	6	Threonine(Thr) 0.6 pmol
7	Alanine(Ala) 0.6 pmol	8	Arginine(Arg) 0.6 pmol	9	Proline(Pro) 0.6 pmol
10	Valine(Val) 0.6 pmol	11	Cystine(Cys-Cys) 0.6 pmol	12	Methionine(Met) 0.6 pmol
13	Ammonium chloride 0.6 pmol	14	Leucine(Leu) 0.6 pmol	15	Isoleucine(Ile) 0.6 pmol
16	Phenylalanine(Phe) 0.6 pmol	17	Lysine(Lys) 0.6 pmol	18	Tyrosine(Tyr) 0.6 pmol

HPLC Conditions

Column CAPCELL PAK C<sub>18</sub> MG S5 : 1.5 mm i.d.×150 mm

Mobile phase (A) 10mmol/L Citrate buffer, 75mmol/L NaClO<sub>4</sub>  
 (B) H<sub>2</sub>O / CH<sub>3</sub>CN = 50 / 50  
 B 5% (0min) - 37% (40min) - 38% (52min) - 100% (60min) - 100% (65min) Gradient

Flow rate            100 µL/min  
Temperature        40 °C  
Detect                FL Ex. 480 nm, Em. 530 nm  
Inj. vol.             1 µL



**AUTHORIZED DISTRIBUTOR**

MZ-Analysentechnik GmbH, Barcelona-Allee 17 • D-55129 Mainz

Tel +49 6131 880 96-0, Fax +49 6131 880 96-20

e-mail: [info@mz-at.de](mailto:info@mz-at.de), [www.mz-at.de](http://www.mz-at.de)